



The Dutch Accreditation Council RvA, by law appointed as the national accreditation body for The Netherlands, hereby declares that accreditation has been granted to:

## **TRESCAL Hengelo B.V. Calibration Laboratory Hengelo**

The organisation has demonstrated to be able to generate technical valid results in a competent way and work according to a management system.

This accreditation is based on an assessment against the requirements as laid down in EN ISO/IEC 17025:2017.

The accreditation covers the activities as specified in the authorized annex bearing the registration number.

The accreditation is valid provided that the organisation continues to meet the requirements.

The accreditation with registration number:

**K 018**

is granted on 15 September 1980

This declaration is valid until

**1 December 2028**

The board of the Dutch Accreditation Council,  
on its behalf,

mr. J.A.W.M. de Haas

A large, stylized handwritten signature in blue ink, which appears to be 'J.A.W.M. de Haas', is written over the printed name.



of **TRESCAL Hengelo B.V.**  
**Calibration Laboratory**

This annex is valid from: **24-04-2025 to 01-12-2028**

Replaces annex dated: **02-10-2024**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
TF 0 0	TIME AND FREQUENCY				
TF 2 0	Relative time				HLO
	Electronic chronometers	24 h	0.1 s / 24 h	Direct measurement	
	Mechanical chronometers	24 h	5 s / 24 h	Direct measurement	
TF 2 1	Time and Frequency				HLO
	10 Hz – 225 MHz		$3 \cdot 10^{-6} \cdot f$	Measure	
TF 2 2	Time interval				HLO
	1 $\mu$ s – 1000 s		$3 \cdot 10^{-6} \cdot t$	Measuring; period applicable to repetitive signals	
	6 min <sup>-1</sup> – 100.000 min <sup>-1</sup>		$4 \cdot 10^{-6} \cdot n$	By comparison with frequency references with n = number of revolutions min <sup>-1</sup>	
TQ 0 0	TORQUE				HLO, OS
TQ 1 2	Torque wrenches	0.1 – 1350 Nm	$1.5 \cdot 10^{-2} \cdot M + 0.5 \cdot r$		

Remarks:

R = reading accuracy of the instrument

Temperature conditions for electrical calibrations is nominal 23 °C; temperature conditions for geometrical and torque calibrations is nominal 20 °C, temperature conditions for pressure and temperature calibrations is nominal 21 °C

$p_e = p - p_{amb}$ :  $p_e$  is overpressure,  $p_{amb}$  is ambient pressure

This list of calibrations is, unless otherwise stated, applicable for calibrations performed inside the laboratory.

<sup>1</sup>  $P_e = P - P_{amb}$ :  $P_e$  is overpressure,  $P_{amb}$  is ambient pressure